· PATENT COOPERATION TREATY

INTERNATIONAL PRELIMINARY REPORT ON PATENTABILITY (Chapter II of the Patent Cooperation Treaty)

(PCT Article 36 and Rule 70)

· · · · · · · · · · · · · · · · · · ·						
Applicant's or agent's file reference 40073PC/GC	FOR FURTHER ACTION	See Form PCT/IPEA/416				
International application No.	International filing date (day/m	nonth/year) Priority date (day/month/year)				
PCT/AU2005/000063	21 January 2005	23 January 2004				
International Patent Classification (IPC) or	national classification and IPC					
Int. Cl.						
1677 577 577 600 600						
A61B 5/154 (2006.01)						
Applicant						
MEDIGARD LIMITED et al						
		ed by this International Preliminary Examining				
Authority under Article 35 and transmit						
2. This REPORT consists of a total of 3	sheets, including this cover shee	xt.				
3. This report is also accompanied by ANN	NEXES, comprising:					
a. X (sent to the applicant and to the	International Bureau) a total of	3 sheets, as follows:				
sheets of the description, claims and/or drawings which have been amended and are the basis for this report and/or sheets containing rectifications authorized by this Authority (see Rule 70.16 and Section 607 of the Administrative Instructions).						
sheets which supersede earlier sheets, but which this Authority considers contain an amendment that goes beyond the disclosure in the international application as filed, as indicated in item 4 of Box No. I and the Supplemental Box.						
b. (sent to the International Bureau only) a total of (indicate type and number of electronic carrier(s)), containing a sequence listing and/or table related thereto, in electronic form only, as indicated in the Supplemental Box Relating to Sequence Listing (see Section 802 of the Administrative Instructions).						
4 This report contains indications relating						
X Box No. I Basis of the repor	t					
Box No. II Priority		•				
Box No. III Non-establishmer	nt of opinion with regard to nove	lty, inventive step and industrial applicability				
Box No. IV Lack of unity of it						
X Box No. V Reasoned statement under Article 35(2) with regard to novelty, inventive step or industrial applicability; citations and explanations supporting such statement						
Box No. VII Certain defects in	the international application					
·	···································	on				
Box No. VIII Certain observations on the international application						
Date of submission of the demand		Date of completion of this report				
15 August 2005	28 April	2006				
Name and mailing address of the IPEA/AU		d Officer				
AUSTRALIAN PATENT OFFICE PO BOX 200, WODEN ACT 2606, AUSTRALIA						
E-mail address: pct@ipaustralia.gov.au	XAVIE	XAVIER GISZ				
Facsimile No. (02) 6285 3929	I Telenhon	ne No. (02) 6283 2064				

INTERNATIONAL PRELIMINARY REPORT ON PATENTABILITY

International application No.

PCT/AU2005/000063

Bo	x No. 1	Basis of the report						
1.	With	regard to the language, this report is based on:						
	X	The international application in the language in which it was filed						
		A translation of the international application into translation furnished for the purposes of:	, which is the language of a					
	•	international search (under Rules 12.3(a) and 23.1 (b))						
		publication of the international application (under Rule 12.4(a))						
		international preliminary examination (Rules 55.2(a) and/or 55.3(a))						
2.	. With regard to the elements of the international application, this report is based on (replacement sheets which have been furnished to the receiving Office in response to an invitation under Article 14 are referred to in this report as "originall filed" and are not annexed to this report):							
		the international application as originally filed/furnished						
	X	the description:	•					
		pages 1-16 as originally filed/furnished						
()		pages* received by this Authority on with the letter of						
		pages* received by this Authority on with the letter of	•					
	X	the claims:						
		pages as originally filed/furnished						
		pages* 17-19 as amended (together with any statement) under Art	icle 19					
		pages* received by this Authority on with the letter of pages* received by this Authority on with the letter of						
	X	the drawings:						
	. لنك	pages 1/8 – 8/8 as originally filed/furnished	·					
		pages* received by this Authority on with the letter of						
		pages* received by this Authority on with the letter of						
		a sequence listing and/or any related table(s) - see Supplemental Box Relating to Sequen	ice Listing.					
3.		The amendments have resulted in the cancellation of:						
		the description, pages						
(·	ì	the claims, Nos.						
٠.		the drawings, sheets/figs	•					
		the sequence listing (specify):						
		any table(s) related to the sequence listing (specify):						
4.	1	This report has been established as if (some of) the amendments annexed to this report ar made, since they have been considered to go beyond the disclosure as filed, as indicated in 70.2(c)).	nd listed below had not been in the Supplemental Box (Rule					
		the description, pages						
	٠.	the claims, Nos.						
		the drawings, sheets/figs						
	•	the sequence listing (specify):						
		any table(s) related to the sequence listing (specify):						
	•							
•	If ite	m 4 applies, some or all of those sheets may be marked "superseded."						

INTERNATIONAL PRELIMINARY REPORT ON PATENTABILITY

International application No.

PCT/AU2005/000063

Box No. V Reasoned statement under Article 35(2) with regard to novelty, inventive step or industrial applicability; citations and explanations supporting such statement

4	Cl	۰
	Statement	•
	омисти	

Novelty (N)	Claims	1-9	YES
	Claims	10	NO
Inventive step (IS)	Claims	1-9	YES
	Claims	10	NO
Industrial applicability (IA)	Claims	1-10	YES
	Claims		NO

2. Citations and explanations (Rule 70.7)

The following documents identified in the International Search Report have been considered for the purposes of this report:

D2.... US 6,572,565

Novelty (N) Claim 10

Claim 10: D2 discloses a blood sampling assembly with a retractable needle. The needle holder and attached needle are held in the cylinder (10) by tabs (20). To retract the needle the health care worker inserts, and rotates the tube (62) relative to cylinder (10) and the needle holder and needle are automatically withdrawn into the tube (62) (column 8 lines 38 to 43).

Inventive Step (IS) Claim 10

Claim 10 also lacks an inventive step for the reasons given above.